

How strong is the influence of temperature for my pH calibration?

Temperature [°C]	pH buffer solution nominal value [pH]		
	4.01	7.00	10.00
5	4.00	7.10	10.23
20	4.00	7.02	10.05
37	4.02	6.98	9.92

DASGIP - Parallel Bioreactor Systems for Unparalleled Results.